



Certificate of Analysis.

Chromium (Cr⁺⁶) ICP standard solution 1000 ppm in in H₂O, Traceable to NIST.

Product Code:	BS13678.
Description:	Chromium (Cr⁺⁶) ICP standard solution 1000 ppm in in H₂O, Traceable to NIST.
Chemical Name:	NA.
Pack Size:	100 ml.
CAS NO.:	NA.
Mol. Weight:	NA.
Mol. Formula:	NA.
Solubility:	Miscible in water.
pH:	NA.
Accuracy:	± 5%.
Storage:	Store at ambient temperature.
LOT NO.:	Sample.
Manufacture Date:	Lot specific.
Expiry Date:	2 years.

Characteristics	Specification	Measured Values
Appearance:	Colorless to Blue with a Green or Purple Cast liquid.	Lot specific.
Odour:	Odourless.	Lot specific.
Composition:	950-1050 ppm @ 25°C.	Lot specific.

Note: - This material is only for laboratory purpose and not for human consumption.

Certification & Traceability:

This product was manufactured, processed and/or certified under a quality management system that is accredited **ISO 17034:2016 and ISO/IEC 17025:2017**. This product was prepared to the certified concentrations shown above by gravimetric methods and analysed using "High Performance ICP-OES" protocol developed by NIST and is directly traceable to **NIST SRM**.

The solution was diluted with filtered (0.22 µm), 18 M-ohm deionized water and stabilized with the appropriate high-purity acid as indicated in the matrix. The balances used in the preparation of this product are calibrated regularly, using a calibration provider that is accredited to **ISO/IEC 17025**. All standard components used in the manufacture of this



product are pre-qualified and verified before use. All volumetric dilutions are performed in Class A calibrated glassware.

Tests were performed for capacity, readability, repeatability and linearity. Secondary verification of the certified concentrations was performed using ICP-OES that was calibrated and/or referenced against **NIST SRM**. This product is manufactured, packaged, stored, and shipped in accordance with good manufacturing practices that is certified to **WHO-GMP**. The uncertainty associated with each certified concentration represents the expanded uncertainty at the 95% confidence level using a coverage factor of **k=2**.

Validity Period:

Briti Scientific standards ensures the accuracy of this product for 2 years from the manufacture date given above, provided the instructions for use are followed.

Quality Certifications:

This product was prepared under a quality management system that is accredited to the following:

ISO 17034:2016 Accredited: Reference Materials Producer, CGI Certificate No. WGMP/22N2594 – General Requirements for the Competence of Reference Material Producers.

ISO 17034 references additional requirements specified in ISO Guide 31 and ISO Guide 35.

ISO/IEC 17025:2017 Accredited: Chemical Testing, CGI Certificate No. UG/23N256- General Requirements for the Competence of Testing and Calibration Laboratories.

ISO 9001:2015 Certified: Quality Management Systems, CGI Certificate No. QMS/23N258.

WHO-GMP Certified: Good Manufacturing Practices, CGI Certificate No. WGMP/22N2594.

Homogeneity:

This product was determined to be homogenous by procedures consistent with the requirements of **ISO 17034** and **ISO Guide 35**. Replicate samples of the finished solution were analysed to confirm its homogeneity, in accordance with internal procedures for the assessment of homogeneity and stability.



Health and Safety Information:

Refer to the Safety Data Sheet (SDS), which can be obtained at www.britisscientific.com.

This certificate shall not be reproduced except in full, without written approval from Briti Scientific. This is a computer generated COA, no stamp or signature is required.



Dr.K. Deepti.
Technical Head.

