



## Certificate of Analysis.

### Viscosity Standard AnStan<sup>®</sup> 5000 cps @25°C.

Product Code:	<b>BS11276.</b>
Discription:	<b>Viscosity Standard AnStan<sup>®</sup> 5000 cps @25°C.</b>
Pack:	<b>500 ml.</b>
CAS No.:	<b>NA.</b>
Mol. Weight:	<b>NA.</b>
Mol. Formula:	<b>NA.</b>
Solubility:	<b>Insoluble in water.</b>
Accuracy:	<b>± 5%.</b>
Storage:	<b>Store at ambient temperature.</b>
LOT NO.:	<b>Sample.</b>
Manufacture Date:	<b>Lot specific.</b>
Expiry Date:	<b>2 years.</b>

Characteristics	Specification	Measured Values
Appearance:	Clear and colourless liquid.	<b>Lot specific.</b>
Viscosity:	4750-5250 cps @25°C.	<b>Lot specific.</b>

Please Note: - This material is only for laboratory purpose and not for human consumption.

#### **Certification & Traceability:**

This product was manufactured, processed and/or certified under a quality management system that is accredited ISO 17034:2016 and ISO/IEC 17025:2017. The certified kinematic viscosities were determined in accordance with ASTM D445. The certified values are based on the primary standard, water at 40°C (ITS-90), with a viscosity of 1.0034 mm<sup>2</sup>/s (cSt) as adopted by NIST, ASTM, and ISO 3666. All values are traceable to reference standards characterized by the master viscometer technique in accordance with ASTM D2162.

All standard components used in the manufacture of this product are pre-qualified and verified before use. All volumetric dilutions are performed in Class A calibrated glassware. The balances used in the preparation of this CRM are calibrated regularly, using a calibration provider that is accredited to ISO/IEC 17025. Tests were performed for capacity, readability, repeatability and linearity.

This product is manufactured, packaged, stored, and shipped in accordance with good manufacturing practices that is certified to WHO-GMP. The uncertainty associated with each certified value represents the expanded uncertainty at the 95% confidence level using a coverage factor of k=2.



**Validity Period:**

Briti Scientific standards ensures the accuracy of this product for 2 Years from the manufacture date given above, provided the instructions for use are followed.

**Quality Certifications:**

This product was prepared under a quality management system that is accredited to the following:

**ISO 17034:2016 Accredited:** Reference Materials Producer, CGI Certificate No. WGMP/22N2594 – General Requirements for the Competence of Reference Material Producers.

**ISO 17034** references additional requirements specified in ISO Guide 31 and ISO Guide 35.

**ISO/IEC 17025:2017 Accredited:** Chemical Testing, CGI Certificate No. UG/23N256- General Requirements for the Competence of Testing and Calibration Laboratories

**ISO 9001:2015 Certified:** Quality Management Systems, CGI Certificate No. QMS/23N258.

**WHO-GMP Certified:** Good Manufacturing Practices, CGI Certificate No. WGMP/22N2594.

**Homogeneity:**

This product was determined to be homogenous by procedures consistent with the requirements of **ISO 17034** and **ISO Guide 35**. Replicate samples of the finished solution were analysed to confirm its homogeneity, in accordance with internal procedures for the assessment of homogeneity and stability.

**Health and Safety Information:**

Refer to the Safety Data Sheet (SDS), which can be obtained at [www.britisscientific.com](http://www.britisscientific.com).

This certificate shall not be reproduced except in full, without written approval from Briti Scientific. This is a computer generated COA, no stamp or signature is required.



**Dr.K. Deepti.**  
**Technical Head.**

