



Certificate of Analysis.

Conductivity Standard 18,000 ± 1% μS/cm @25°C, Traceable to NIST.

Product Code:	BSCON18K.
Description:	Conductivity Standard 18,000 ± 1% μS/cm @25°C, Traceable to NIST.
Pack:	500 ml.
Chemical Name:	NA.
CAS No.:	NA.
Mol. Weight:	NA.
Mol. Formula:	NA.
Solubility:	Miscible in water.
pH:	NA.
Accuracy:	± 1% μS/cm.
Storage:	Store at ambient temperature.
Batch No:	Sample.
Manufacture Date:	Lot specific.
Expiry Date:	1 year.

Characteristics	Specification	Measured Values
Appearance:	Clear colourless liquid.	Lot specific.
Odour:	Odourless.	Lot specific.
Conductivity:	17,820 – 18,180 μS/cm @25°C.	Lot specific.

Please Note: - This material is only for laboratory purpose and not for human consumption.

This product is prepared using high quality salts and deionized water. The results reported are of a sample taken from the above-mentioned lot during manufacture and measurements are in comparison to standard prepared from **NIST SRM 999c**.

Certification & Traceability:

This product was manufactured, processed and/or certified under a quality management system that is accredited **ISO 17034:2016 and ISO/IEC 17025:2017**. This product was prepared to the certified concentrations shown above by gravimetric methods and analysed using a calibrated conductivity meter protocol and is directly traceable to **NIST SRM**.

The solution was diluted with filtered (0.22 μm), 18 M-ohm deionized water. The balances used in the preparation of this product are calibrated regularly, using a calibration provider that is accredited to **ISO/IEC 17025**. All standard components used in the manufacture of this product are pre-qualified and verified before use. All volumetric dilutions are performed in Class A calibrated glassware.



USA: #1004, Boston, Massachusetts, 02116, United States. **India:** Jeedimetla, Hyderabad-500 055, Telangana.

www.britisscientific.com

Tests were performed for capacity, readability, repeatability and linearity. This product is manufactured, packaged, stored, and shipped in accordance with good manufacturing practices that is certified to **WHO-GMP**. The uncertainty associated with each certified concentration represents the expanded uncertainty at the 95% confidence level using a coverage factor of **k=2**.

Validity Period:

Briti Scientific standards ensures the accuracy of this product for 12 months from the manufacture date given above, provided the instructions for use are followed.

Quality Certifications:

This product was prepared under a quality management system that is Complies with the following:

ISO 17034:2016: Reference Materials Producer, CGI Certificate No. WGMP/22N2594 – General Requirements for the Competence of Reference Material Producers.

ISO 17034 references additional requirements specified in ISO Guide 31 and ISO Guide 35.

ISO/IEC 17025:2017: Chemical Testing, CGI Certificate No. UG/23N256- General Requirements for the Competence of Testing and Calibration Laboratories.

ISO 9001:2015 Certified: Quality Management Systems, CGI Certificate No. QMS/23N258.

WHO-GMP Certified: Good Manufacturing Practices, CGI Certificate No. WGMP/22N2594.

Homogeneity:

This product was determined to be homogenous by procedures consistent with the requirements of **ISO 17034** and **ISO Guide 35**. Replicate samples of the finished solution were analysed to confirm its homogeneity, in accordance with internal procedures for the assessment of homogeneity and stability.

Health and Safety Information:

Refer to the Safety Data Sheet (SDS), which can be obtained at www.britisscientific.com.

This certificate shall not be reproduced except in full, without written approval from Briti Scientific. This is a computer generated COA, no stamp or signature is required

N. Bharathi.
Quality Control.

N.S. Mallika.
Quality Assurance.



USA: #1004, Boston, Massachusetts, 02116, United States. **India:** Jeedimetla, Hyderabad-500 055, Telangana.

www.britisscientific.com