



## Certificate of Analysis.

**Conductivity Standard 100000 ± 1% µS/cm @ 25°C, Traceable to NIST.**

Product Code:	<b>BSCON100K.</b>
Description:	<b>Conductivity Standard 100000 ± 1% µS/cm @ 25°C, Traceable to NIST.</b>
Pack:	<b>500 ml.</b>
Chemical Name:	<b>NA.</b>
CAS No.:	<b>NA.</b>
Mol. Weight:	<b>NA.</b>
Mol. Formula:	<b>NA.</b>
Solubility:	<b>Miscible in water.</b>
Accuracy:	<b>± 1% µS/cm.</b>
Storage:	<b>Store at ambient temperature.</b>
LOT NO.:	<b>Sample.</b>
Manufacture Date:	<b>Lot specific.</b>
Expiry Date:	<b>1 year.</b>

Characteristics	Specification	Measured Values
Appearance:	Clear and colourless liquid.	<b>Lot specific.</b>
Odour:	Odourless.	<b>Lot specific.</b>
Conductivity:	99000-101000 µS/cm @25°C.	<b>Lot specific.</b>

Please Note: - This material is only for laboratory purpose and not for human consumption.

This product is prepared using high quality salts and deionized water. The results reported are of a sample taken from the above-mentioned lot during manufacture and measurements are in comparison to standard prepared from **NIST SRM 999c**. The uncertainty of the mean specific value does not exceed ±1% at 95% confidence level (k=2).

All standard components used in the manufacture of this product are pre-qualified and verified before use. All glassware used are of class A and have been qualified and verified before use. Tests were performed for capacity, readability, repeatability and linearity. This product meets all requirements defined by the appropriate specifications. This product is manufactured, packaged, stored, and shipped in accordance with good manufacturing practices.

This is a computer-generated COA; no stamp or signature is required This certificate shall not be reproduced except in full, without written approval from Briti Scientific.



**Dr.K. Deepti.**  
**Technical Head.**

