

Certificate of Analysis.

Silodosin secondary standard, certified reference material (CRM).


Product code:	BS16050.
Description:	Silodosin secondary standard, certified reference material (CRM).
Synonym:	(R)-1-(3-Hydroxypropyl)-5-[2-[[2-(2,2,2-trifluoroethoxy)phenoxy)ethyl]amino]propyl]indoline-7-carboxamide.
Pack:	500 mg.
CAS NO.:	160970-54-7.
Mol. Weight:	495.54 g/mol.
Mol. Formula:	C₂₅H₃₂F₃N₃O₄.
MDL Number:	MFCD00930170.
Melting point:	107 °C.
Boiling Point:	601.4°C at 760 mmHg.
Density:	1.249 g/cm³.
Solubility:	Soluble in methanol.
Storage:	Store at 2-8°C Refrigerator.
Batch No:	Sample.
Manufacture Date:	Lot specific.
Expiry Date:	3 years.

Test	Specification	Measured Values
Appearance:	White to Light yellow powder to crystal.	Lot specific.
Identification by NMR:	Confirms to the structure.	Lot specific.
Identification by MASS:	Confirms to the structure.	Lot specific.
Identification by IR:	Confirms to the structure.	Lot specific.
Purity by HPLC:	≥99.5%.	Lot specific.

Comparative Purity:

Test	In-house LOT (Lot specific)	USP LOT (Lot specific)
Purity by HPLC:	Lot specific.	Lot specific.

Please Note: - This material is only for laboratory purpose and not for human consumption. This is a computer generated COA, no stamp or signature is required.



Certification & Traceability:

This product was manufactured, processed and/or certified under a quality management system that complies with ISO 17034:2016 and ISO/IEC 17025:2017. This standard has been gravimetrically



USA: #1004, Boston, Massachusetts, 02116, United States. **India:** Jeedimetla, Hyderabad-500 055, Telangana.

www.britishtscientific.com

prepared Using balances that have been fully qualified and calibrated to ISO 17025 requirements. All calibrations utilize NIST traceable weights which are calibrated externally by a qualified ISO 17025 accredited calibration laboratory to NIST standards. Qualification of each balance includes the assignment of a minimum weighing by a qualified and ISO 17025 accredited calibration vendor taking into consideration the balance and installed environmental conditions to ensure compliance with USP tolerances of NMT 0.10% relative error.

All standard components used in the manufacture of this product are pre-qualified and verified before use. Tests were performed for capacity, readability, repeatability and linearity.

This product is manufactured, packaged, stored, and shipped in accordance with good manufacturing practices that is certified to WHO-GMP. The uncertainty associated with each certified value represents the expanded uncertainty at the 95% confidence level using a coverage factor of $k=2$.

Instructions for handling and correct use: The internal pressure of the container may be slightly different from the atmospheric pressure at the users location. Open slowly and carefully to avoid dispersion of the material.

Validity Period:

Briti Scientific standards ensures the accuracy of this product for 3 years from the manufacture date given above, provided the instructions for use are followed.

Quality Certifications:

This product was prepared under a quality management system that is Complies with the following:

ISO 17034:2016: Reference Materials Producer, CGI Certificate No. WGMP/22N2594 – General Requirements for the Competence of Reference Material Producers.

ISO 17034 references additional requirements specified in ISO Guide 31 and ISO Guide 35.

ISO/IEC 17025:2017: Chemical Testing, CGI Certificate No. UG/23N256- General Requirements for the Competence of Testing and Calibration Laboratories

ISO 9001:2015 Certified: Quality Management Systems, CGI Certificate No. QMS/23N258.

WHO-GMP Certified: Good Manufacturing Practices, CGI Certificate No. WGMP/22N2594.

Homogeneity:

This product was determined to be homogenous by procedures consistent with the requirements of ISO 17034 and ISO Guide 35. Replicate samples of the finished solution were analysed to confirm its homogeneity, in accordance with internal procedures for the assessment of homogeneity and stability.

Health and Safety Information:



USA: #1004, Boston, Massachusetts, 02116, United States. India: Jeedimetla, Hyderabad-500 055, Telangana.

www.britisscientific.com

Refer to the Safety Data Sheet (SDS), which can be obtained at www.britishtscientific.com.

Disclaimer: The purchaser is required to determine the suitability of this product for any particular application. Briti scientific makes no warranty of any kind, express or implied, other than its products meet all quality control standards set by Briti scientific. We do not guarantee that the product can be used for any particular application.

This certificate shall not be reproduced except in full, without written approval from Briti Scientific. This is a computer generated COA, no stamp or signature is required.



USA: #1004, Boston, Massachusetts, 02116, United States. **India:** Jeedimetla, Hyderabad-500 055, Telangana.
www.britishtscientific.com